

SuperGuard™

MEDICAL EXAMINATION GLOVES

• **HIGH NEWTON
BLUPLE**



• **POWDER FREE
NITRILE**

SUPERMAX GLOVE MANUFACTURING CORPORATION SDN BHD is a leading Manufacturer of high quality gloves, exporting to over 146 countries worldwide.

Supermax's *SuperGuard* range of Nitrile Gloves are used extensively in manufacturing and production facilities in processes, which require additional protection for workers and their products.

PRODUCT INFORMATION:

Product Type	: Non-Sterile Powder Free Medical Examination Gloves (5 mil Thickness)
Material	: Nitrile
Colour	: Blue
Usage	: For Single Use Only
Design	: Ambidextrous, Textured, Beaded Cuff
Powder Free Residue	: Max 2mg/glove (Typical: 0.2mg/glove)

SUPER GUARD GLOVES ARE IDEAL FOR USE IN :

- Chemical/Electronic Industries.
- Biotechnology Industries.
- Laboratories/Pharmaceutical Industries.
- Health Care Institutions.
- Oncology Units
- Special Packing Industries.

STANDARDS :

- ASTM D 6319, EN 455 Part 1,2,3,4.
- AQL 1.5 Quality System Standards and Gold Seal Standard.
- CE Certified Company.

QUALITY ASSURANCE:

- ISO 9001 Certified by SGS (UK) and SIRIM (Malaysia).
- ISO 13485 Certified by SGS (UK).
- CMDCAS ISO 13485 Certification.

Manufactured by:

SUPERMAX GLOVE MANUFACTURING SDN.BHD.
Malaysia



Imported by:

PAMIDI RSB IMPEX

Hyderabad-500 081. India.

Customer care No. +91 406 455 6674.

Email : pamidiimpex@gmail.com.



DIMENSIONS				
Size	Palm Width (mm)		Length (mm)	
	ASTM D 6319	SUPERMAX	ASTM D 6319	SUPERMAX
XS	70 ± 10	70 - 79	Min.220	Min.240
S	80 ± 10	80 - 89	Min.220	Min.240
M	95 ± 10	90 - 99	Min.230	Min.240
L	110 ± 10	100 - 109	Min.230	Min.240
XL	120 ± 10	110 - 119	Min.230	Min.240

THICKNESS		
Location of Thickness Measurement	Single Wall (mm)	
	ASTM D 6319	SUPERMAX
Finger	Min.0.05	Min.0.13
Palm	Min.0.05	Min.0.10
Cuff	N/A	Min.0.08

PHYSICAL PROPERTIES				
Test	Before Ageing		After Ageing	
	ASTM D 6319	SUPER MAX	ASTM D6319	SUPER MAX
Tensile Strength(Mpa)	Min.14	Min.14	Min.14	Min.14
Elongation (%)	Min.500	Min.500	Min.400	Min.400
Force at Break (N)	N/A	Min.9	N/A	Min.6

INSPECTION CRITERIA (Sampling & Inspection ISO 2859)		
Test	Inspection Level	AQL
Watertight Test	G1	1.5
Major Defects	G1	2.5
Minor Defects	G1	4
Dimensions	S2	4
Physical properties	S2	4

PACKING SPECIFICATIONS	
Inner Box (Qty)	100 ± 2 Pieces
Outer Carton (Qty)	10 Inner Boxes

SuperGuard™

- HIGH NEWTON BLUPL • POWDER FREE NITRILE
- ## Medical Examination Gloves

Marketed by:

PAMIDI RSB Build-Tech Pvt. Ltd.

104, Xeno Summit Apartment, Jaihind Enclave, Madhapur,
Hyderabad-500 081. India.

Ph: 040 40153125, Mobile: +91 91005 50059.

Email : pamidigloves@gmail.com.